

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032712.D
 Acq On : 10 Nov 2022 07:13
 Operator : JC/MD
 Sample : N5509-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-12

Quant Time: Nov 10 08:09:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120051	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88028	52.772	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.540%
35) Dibromofluoromethane	5.385	113	67634	46.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.980%
50) Toluene-d8	8.647	98	253905	50.191	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.380%
62) 4-Bromofluorobenzene	11.079	95	100864	51.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.080%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	16107	26.567	ug/l	# 88
20) Methylene Chloride	2.794	84	6885	4.953	ug/l	97
37) Ethyl Acetate	4.714	43	15224	7.219	ug/l	98
43) Isopropyl Acetate	6.342	43	73242	22.620	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032712.D
Acq On : 10 Nov 2022 07:13
Operator : JC/MD
Sample : N5509-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-12

Quant Time: Nov 10 08:09:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

