

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032715.D
 Acq On : 10 Nov 2022 08:22
 Operator : JC/MD
 Sample : N5511-03ME
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-05-110722ME

Quant Time: Nov 10 08:48:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221806	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	123326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91618	49.254	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.500%
35) Dibromofluoromethane	5.391	113	67697	42.788	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.580%
50) Toluene-d8	8.653	98	301719	54.843	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.680%
62) 4-Bromofluorobenzene	11.085	95	128634	59.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.720%

Target Compounds						Qvalue
31) Cyclohexane	5.501	56	192985	85.567	ug/l #	48
39) Methylcyclohexane	7.379	83	260680	100.001	ug/l	94
52) Toluene	8.720	92	21696	5.334	ug/l	99
67) Ethyl Benzene	10.195	91	1174218	128.800	ug/l	99
68) m/p-Xylenes	10.305	106	3214017	907.903	ug/l	90
69) o-Xylene	10.646	106	784278	227.863	ug/l	98
73) Isopropylbenzene	10.963	105	609668	62.386	ug/l	100
78) n-propylbenzene	11.311	91	1611324	143.110	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	2418128	289.518	ug/l	98
84) 1,2,4-Trimethylbenzene	11.762	105	7144017	859.984	ug/l	90
85) sec-Butylbenzene	11.896	105	340090	33.895	ug/l	93
86) p-Isopropyltoluene	12.012	119	185706	21.812	ug/l	99
89) n-Butylbenzene	12.335	91	1161664	156.837	ug/l #	41
95) Naphthalene	13.780	128	1207150	124.683	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032715.D
Acq On : 10 Nov 2022 08:22
Operator : JC/MD
Sample : N5511-03ME
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-05-110722ME

Quant Time: Nov 10 08:48:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

