

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110922\
 Data File : VX032697.D
 Acq On : 10 Nov 2022 01:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 03:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	50.000	49.002	2.0	75	0.00
3 P	Chloromethane	50.000	44.592	10.8	71	0.00
4 C	Vinyl Chloride	50.000	48.172	3.7#	75	0.00
5 T	Bromomethane	50.000	39.309	21.4	62	0.01
6 T	Chloroethane	50.000	65.594	-31.2#	113	0.00
7 T	Trichlorofluoromethane	50.000	50.751	-1.5	83	0.00
8 T	Diethyl Ether	50.000	50.725	-1.5	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.474	-2.9	86	0.00
10 T	Methyl Iodide	50.000	41.381	17.2	67	0.00
11 T	Tert butyl alcohol	250.000	249.652	0.1	83	0.04
12 CM	1,1-Dichloroethene	50.000	49.846	0.3#	82	0.00
13 T	Acrolein	250.000	250.039	-0.0	85	0.00
14 T	Allyl chloride	50.000	51.754	-3.5	84	0.00
15 T	Acrylonitrile	250.000	242.477	3.0	83	0.00
16 T	Acetone	250.000	232.537	7.0	83	0.01
17 T	Carbon Disulfide	50.000	44.023	12.0	70	0.00
18 T	Methyl Acetate	50.000	53.473	-6.9	89	0.00
19 T	Methyl tert-butyl Ether	50.000	54.527	-9.1	92	0.00
20 T	Methylene Chloride	50.000	49.860	0.3	88	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.234	-0.5	85	0.00
22 T	Diisopropyl ether	50.000	55.731	-11.5	94	0.00
23 T	Vinyl Acetate	250.000	263.231	-5.3	87	0.00
24 P	1,1-Dichloroethane	50.000	53.717	-7.4	90	0.00
25 T	2-Butanone	250.000	234.680	6.1	80	0.00
26 T	2,2-Dichloropropane	50.000	49.863	0.3	82	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.014	-8.0	89	0.00
28 T	Bromochloromethane	50.000	48.957	2.1	86	0.00
29 T	Tetrahydrofuran	250.000	236.730	5.3	79	0.00
30 C	Chloroform	50.000	55.865	-11.7#	94	0.00
31 T	Cyclohexane	50.000	50.491	-1.0	82	0.00
32 T	1,1,1-Trichloroethane	50.000	55.332	-10.7	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.569	-3.1	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	84	0.00
35 S	Dibromofluoromethane	50.000	51.648	-3.3	89	0.00
36 T	1,1-Dichloropropene	50.000	51.239	-2.5	86	0.00
37 T	Ethyl Acetate	50.000	45.884	8.2	82	0.00
38 T	Carbon Tetrachloride	50.000	54.054	-8.1	89	0.00
39 T	Methylcyclohexane	50.000	48.788	2.4	79	0.00
40 TM	Benzene	50.000	52.170	-4.3	87	0.00
41 T	Methacrylonitrile	50.000	52.399	-4.8	88	0.00
42 TM	1,2-Dichloroethane	50.000	54.324	-8.6	91	0.00
43 T	Isopropyl Acetate	50.000	51.372	-2.7	85	0.00
44 TM	Trichloroethene	50.000	51.298	-2.6	87	0.00
45 C	1,2-Dichloropropane	50.000	53.485	-7.0#	90	0.00
46 T	Dibromomethane	50.000	52.887	-5.8	90	0.00
47 T	Bromodichloromethane	50.000	55.464	-10.9	93	0.00
48 T	Methyl methacrylate	50.000	51.476	-3.0	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032697.D
 Acq On : 10 Nov 2022 01:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 03:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	891.148	10.9	73	0.04
50 S	Toluene-d8	50.000	49.720	0.6	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	249.502	0.2	83	0.00
52 CM	Toluene	50.000	53.655	-7.3#	89	0.00
53 T	t-1,3-Dichloropropene	50.000	56.418	-12.8	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.310	-8.6	89	0.00
55 T	1,1,2-Trichloroethane	50.000	53.180	-6.4	92	0.00
56 T	Ethyl methacrylate	50.000	55.045	-10.1	88	0.00
57 T	1,3-Dichloropropane	50.000	53.695	-7.4	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.681	-3.1	83	0.00
59 T	2-Hexanone	250.000	247.558	1.0	81	0.00
60 T	Dibromochloromethane	50.000	54.379	-8.8	90	0.00
61 T	1,2-Dibromoethane	50.000	54.043	-8.1	90	0.00
62 S	4-Bromofluorobenzene	50.000	53.501	-7.0	89	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	85	0.00
64 T	Tetrachloroethene	50.000	51.234	-2.5	89	0.00
65 PM	Chlorobenzene	50.000	52.012	-4.0	91	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.793	-7.6	93	0.00
67 C	Ethyl Benzene	50.000	53.041	-6.1#	90	0.00
68 T	m/p-Xylenes	100.000	105.063	-5.1	89	0.00
69 T	o-Xylene	50.000	53.193	-6.4	90	0.00
70 T	Styrene	50.000	54.326	-8.7	91	0.00
71 P	Bromoform	50.000	53.640	-7.3	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	84	0.00
73 T	Isopropylbenzene	50.000	50.927	-1.9	90	0.00
74 T	N-ethyl acetate	50.000	51.906	-3.8	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.134	1.7	88	0.00
76 T	1,2,3-Trichloropropane	50.000	50.807	-1.6	87	0.00
77 T	Bromobenzene	50.000	51.808	-3.6	93	0.00
78 T	n-propylbenzene	50.000	50.780	-1.6	88	0.00
79 T	2-Chlorotoluene	50.000	51.884	-3.8	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.622	-3.2	88	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.923	4.2	80	0.00
82 T	4-Chlorotoluene	50.000	51.746	-3.5	90	0.00
83 T	tert-Butylbenzene	50.000	52.160	-4.3	90	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.173	-4.3	89	0.00
85 T	sec-Butylbenzene	50.000	51.506	-3.0	87	0.00
86 T	p-Isopropyltoluene	50.000	51.419	-2.8	86	0.00
87 T	1,3-Dichlorobenzene	50.000	52.102	-4.2	92	0.00
88 T	1,4-Dichlorobenzene	50.000	49.887	0.2	89	0.00
89 T	n-Butylbenzene	50.000	50.399	-0.8	84	0.00
90 T	Hexachloroethane	50.000	51.173	-2.3	87	0.00
91 T	1,2-Dichlorobenzene	50.000	52.519	-5.0	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.780	6.4	79	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.595	0.8	89	0.00
94 T	Hexachlorobutadiene	50.000	47.738	4.5	86	0.00
95 T	Naphthalene	50.000	48.769	2.5	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032697.D
Acq On : 10 Nov 2022 01:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 10 03:08:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.931	2.1	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6