


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032696.D
Acq On : 10 Nov 2022 00:41
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Title : SW846 8260
Last Update : Thu Oct 27 08:28:06 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	6879	PASS
75	95	30	60	56.3	18299	PASS
95	95	100	100	100.0	32507	PASS
96	95	5	9	6.9	2227	PASS
173	174	0.00	2	0.9	241	PASS
174	95	50	100	78.4	25485	PASS
175	174	5	9	7.5	1915	PASS
176	174	95	101	96.3	24551	PASS
177	176	5	9	6.7	1651	PASS