

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038676.D
 Acq On : 09 Nov 2023 13:32
 Operator : JC/MD
 Sample : 05222-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-27-20231101

Quant Time: Nov 10 03:38:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	309622	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	506957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	528149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	322949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175319	45.482	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.960%
35) Dibromofluoromethane	5.391	113	170757	46.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.647	98	580987	44.467	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.940%#
62) 4-Bromofluorobenzene	11.079	95	272298	53.753	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.500%
Target Compounds						
					Qvalue	
40) Benzene	6.043	78	8473	0.649	ug/l	# 83
52) Toluene	8.720	92	15404	1.800	ug/l	98
68) m/p-Xylenes	10.299	106	4587	0.636	ug/l	98
69) o-Xylene	10.646	106	1798	0.256	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038676.D
Acq On : 09 Nov 2023 13:32
Operator : JC/MD
Sample : 05222-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-27-20231101

Quant Time: Nov 10 03:38:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

