

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038684.D
 Acq On : 09 Nov 2023 16:38
 Operator : JC/MD
 Sample : 05241-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-TB-7

Quant Time: Nov 10 03:41:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	256325	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	420025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	435754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149971	46.996	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.000%
35) Dibromofluoromethane	5.385	113	139010	45.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	8.647	98	497222	45.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.860%#
62) 4-Bromofluorobenzene	11.079	95	212528	50.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038684.D
Acq On : 09 Nov 2023 16:38
Operator : JC/MD
Sample : 05241-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-TB-7

Quant Time: Nov 10 03:41:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

