

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038687.D
 Acq On : 09 Nov 2023 17:47
 Operator : JC/MD
 Sample : 05241-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-13-20231102

Quant Time: Nov 10 03:42:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	248544	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	417760	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	437356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	256447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148840	48.102	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.200%		
35) Dibromofluoromethane	5.385	113	137487	45.694	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.380%		
50) Toluene-d8	8.647	98	497353	46.193	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.380%		
62) 4-Bromofluorobenzene	11.079	95	213897	51.240	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.480%		
Target Compounds						
32) 1,1,1-Trichloroethane	5.385	97	18958	4.432	ug/l	97
44) Trichloroethene	7.129	130	484778	144.732	ug/l	91
64) Tetrachloroethene	9.275	164	1498	0.458	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038687.D
Acq On : 09 Nov 2023 17:47
Operator : JC/MD
Sample : 05241-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-13-20231102

Quant Time: Nov 10 03:42:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

