

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038688.D
 Acq On : 09 Nov 2023 18:10
 Operator : JC/MD
 Sample : 05241-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-FD-5-20231102

Quant Time: Nov 10 03:43:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	250990	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	423848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	433498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	246746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148676	47.580	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.160%
35) Dibromofluoromethane	5.391	113	139541	45.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.420%
50) Toluene-d8	8.647	98	498614	45.645	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.300%#
62) 4-Bromofluorobenzene	11.079	95	210584	49.722	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.440%
Target Compounds						
					Qvalue	
32) 1,1,1-Trichloroethane	5.385	97	18716	4.333	ug/l	96
38) Carbon Tetrachloride	5.684	117	1190	0.281	ug/l	85
44) Trichloroethene	7.129	130	476649	140.261	ug/l	90
64) Tetrachloroethene	9.275	164	1449	0.447	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038688.D
Acq On : 09 Nov 2023 18:10
Operator : JC/MD
Sample : 05241-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-FD-5-20231102

Quant Time: Nov 10 03:43:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

