

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038690.D
 Acq On : 09 Nov 2023 18:56
 Operator : JC/MD
 Sample : 05241-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-GEP-1-20231102

Quant Time: Nov 10 03:44:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	260377	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	435902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	455839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	262579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152846	47.151	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.300%		
35) Dibromofluoromethane	5.391	113	144165	45.920	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.840%		
50) Toluene-d8	8.647	98	518323	46.137	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.280%		
62) 4-Bromofluorobenzene	11.079	95	220105	50.533	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.060%		
Target Compounds						
16) Acetone	2.374	43	3042	2.228	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	21657	4.833	ug/l	95
38) Carbon Tetrachloride	5.678	117	1933	0.443	ug/l #	81
44) Trichloroethene	7.129	130	583885	167.065	ug/l	92
64) Tetrachloroethene	9.275	164	2646	0.775	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038690.D
Acq On : 09 Nov 2023 18:56
Operator : JC/MD
Sample : 05241-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-GEP-1-20231102

Quant Time: Nov 10 03:44:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

