

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038691.D
 Acq On : 09 Nov 2023 19:20
 Operator : JC/MD
 Sample : 05241-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-24-20231102

Quant Time: Nov 10 03:44:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270665	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	447244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	467106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	264399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156759	46.520	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.040%
35) Dibromofluoromethane	5.385	113	145070	45.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.080%
50) Toluene-d8	8.647	98	534614	46.381	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.760%
62) 4-Bromofluorobenzene	11.079	95	226087	50.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.180%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	90978	36.017	ug/l	# 76
16) Acetone	2.380	43	4341	3.059	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	3274	1.191	ug/l	# 86
27) cis-1,2-Dichloroethene	4.495	96	16291	5.130	ug/l	77
44) Trichloroethene	7.123	130	712811	198.782	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038691.D
Acq On : 09 Nov 2023 19:20
Operator : JC/MD
Sample : 05241-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-24-20231102

Quant Time: Nov 10 03:44:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

