

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038701.D
 Acq On : 10 Nov 2023 03:54
 Operator : JC/MD
 Sample : 05241-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-33-20231102

Quant Time: Nov 10 05:28:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	272996	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	449652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	468195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156102	45.930	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.860%
35) Dibromofluoromethane	5.385	113	146585	45.263	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	8.647	98	535509	46.209	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.420%
62) 4-Bromofluorobenzene	11.079	95	229459	51.069	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.140%
Target Compounds						
44) Trichloroethene	7.129	130	619952	171.961	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038701.D
Acq On : 10 Nov 2023 03:54
Operator : JC/MD
Sample : 05241-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-33-20231102

Quant Time: Nov 10 05:28:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

