

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038702.D
 Acq On : 10 Nov 2023 04:17
 Operator : JC/MD
 Sample : 05241-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-32-20231102

Quant Time: Nov 10 05:28:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	281762	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	461866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	473606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	266195	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161178	45.948	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.900%
35) Dibromofluoromethane	5.385	113	149476	44.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.860%
50) Toluene-d8	8.647	98	548889	46.111	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.220%
62) 4-Bromofluorobenzene	11.079	95	231520	50.165	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.340%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3689	2.497	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	4689	1.418	ug/l	70
44) Trichloroethene	7.129	130	461298	124.570	ug/l	91
52) Toluene	8.720	92	3095	0.397	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038702.D
Acq On : 10 Nov 2023 04:17
Operator : JC/MD
Sample : 05241-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-32-20231102

Quant Time: Nov 10 05:28:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

