

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038704.D
 Acq On : 10 Nov 2023 05:03
 Operator : JC/MD
 Sample : 05241-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-26-20231102

Quant Time: Nov 10 06:21:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	266339	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	440569	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	459104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	263587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153676	46.346	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.700%
35) Dibromofluoromethane	5.385	113	143688	45.283	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.560%
50) Toluene-d8	8.647	98	525416	46.273	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.540%
62) 4-Bromofluorobenzene	11.079	95	224013	50.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.780%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	3576	2.561	ug/l	# 84
44) Trichloroethene	7.135	130	2718	0.769	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038704.D
Acq On : 10 Nov 2023 05:03
Operator : JC/MD
Sample : 05241-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-26-20231102

Quant Time: Nov 10 06:21:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

