

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038706.D
 Acq On : 10 Nov 2023 05:49
 Operator : JC/MD
 Sample : 05241-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-35-20231102

Quant Time: Nov 10 06:21:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	271647	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	449174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	478127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	157026	46.431	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.860%
35) Dibromofluoromethane	5.391	113	148822	46.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.000%
50) Toluene-d8	8.647	98	537119	46.398	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	232827	51.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.740%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	146714	40.738	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038706.D
Acq On : 10 Nov 2023 05:49
Operator : JC/MD
Sample : 05241-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-35-20231102

Quant Time: Nov 10 06:21:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

