

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
 Data File : VX038707.D
 Acq On : 10 Nov 2023 06:12
 Operator : JC/MD
 Sample : 05241-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EVAL-01D-20231102

Quant Time: Nov 10 16:42:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	257546	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	433092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	449330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150342	46.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.780%
35) Dibromofluoromethane	5.385	113	143054	45.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.720%
50) Toluene-d8	8.647	98	513214	45.979	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.960%#
62) 4-Bromofluorobenzene	11.079	95	218050	50.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.780%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	3252	1.353	ug/l	# 76
16) Acetone	2.380	43	3560	2.636	ug/l	96
21) trans-1,2-Dichloroethene	3.099	96	1398	0.534	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	13087	4.331	ug/l	75
44) Trichloroethene	7.129	130	1658495	477.619	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038707.D
Acq On : 10 Nov 2023 06:12
Operator : JC/MD
Sample : 05241-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EVAL-01D-20231102

Quant Time: Nov 10 16:42:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

