

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110923\  
 Data File : VX038681.D  
 Acq On : 09 Nov 2023 15:28  
 Operator : JC/MD  
 Sample : 05241-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FSND-TB-5

Quant Time: Nov 10 03:40:30 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X110723W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 08 03:19:07 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 257197   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 426636   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 438520   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 252284   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 150207   | 46.910   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 93.820%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 139639   | 45.444   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 90.880%  |
| 50) Toluene-d8              | 8.647  | 98    | 501664   | 45.624   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 91.240%# |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 216202   | 50.715   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 101.420% |


|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110923\  
Data File : VX038681.D  
Acq On : 09 Nov 2023 15:28  
Operator : JC/MD  
Sample : 05241-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-TB-5

Quant Time: Nov 10 03:40:30 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X110723W.M  
Quant Title : SW846 8260  
QLast Update : Wed Nov 08 03:19:07 2023  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 733  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28




Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-TB-5


Tgt Ion:168 Resp: 257197  
Ion Ratio Lower Upper  
168 100  
99 53.2 56.6 85.0#



#33  
1,2-Dichloroethane-d4  
Concen: 46.910 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28

Tgt Ion: 65 Resp: 150207  
Ion Ratio Lower Upper  
65 100  
67 52.3 0.0 100.2









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 911  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28




Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-TB-5


Tgt Ion:114 Resp: 426636  
Ion Ratio Lower Upper  
114 100  
63 17.3 0.0 45.0  
88 14.7 0.0 36.6



#35  
Dibromofluoromethane  
Concen: 45.444 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28

Tgt Ion:113 Resp: 139639  
Ion Ratio Lower Upper  
113 100  
111 102.5 82.9 124.3  
192 21.2 13.3 19.9#









#50  
Toluene-d8  
Concen: 45.624 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28




Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-TB-5


Tgt Ion: 98 Resp: 501664  
Ion Ratio Lower Upper  
98 100  
100 64.8 53.1 79.7



#62  
4-Bromofluorobenzene  
Concen: 50.715 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28

Tgt Ion: 95 Resp: 216202  
Ion Ratio Lower Upper  
95 100  
174 87.0 0.0 135.6  
176 83.1 0.0 131.0









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28

Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-TB-5

Tgt Ion:117 Resp: 438520  
Ion Ratio Lower Upper  
117 100  
82 49.6 46.2 69.4  
119 32.2 25.5 38.3



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX038681.D  
Acq: 09 Nov 2023 15:28

Tgt Ion:152 Resp: 252284  
Ion Ratio Lower Upper  
152 100  
115 57.7 43.3 129.9  
150 156.3 0.0 345.0

