



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110923\
Data File : VX038668.D
Acq On : 09 Nov 2023 08:52
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Title : SW846 8260
Last Update : Wed Nov 08 03:19:07 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.3	8573	PASS
75	95	30	60	48.6	25517	PASS
95	95	100	100	100.0	52515	PASS
96	95	5	9	7.0	3669	PASS
173	174	0.00	2	0.8	350	PASS
174	95	50	100	85.5	44880	PASS
175	174	5	9	7.4	3328	PASS
176	174	95	101	95.5	42867	PASS
177	176	5	9	6.7	2863	PASS