

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032719.D
 Acq On : 10 Nov 2022 11:41
 Operator : JC/MD
 Sample : VX1110MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1110MBL01

Quant Time: Nov 10 12:09:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	209258	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88894	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89529	51.106	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.220%
35) Dibromofluoromethane	5.385	113	68673	46.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	253554	48.851	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.700%
62) 4-Bromofluorobenzene	11.079	95	99637	49.144	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032719.D
Acq On : 10 Nov 2022 11:41
Operator : JC/MD
Sample : VX1110MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1110MBL01

Quant Time: Nov 10 12:09:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

