

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032735.D
 Acq On : 10 Nov 2022 17:50
 Operator : JC/MD
 Sample : N5558-10DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 08BR-20221109DL

Quant Time: Nov 11 00:37:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	179337	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76871	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76970	51.339	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.680%
35) Dibromofluoromethane	5.391	113	59683	46.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.653	98	220010	49.461	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	85579	49.253	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.500%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1166	1.007	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	12719	9.342	ug/l	87
44) Trichloroethene	7.128	130	65813	46.377	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032735.D
Acq On : 10 Nov 2022 17:50
Operator : JC/MD
Sample : N5558-10DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
08BR-20221109DL

Quant Time: Nov 11 00:37:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

