

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032737.D
 Acq On : 10 Nov 2022 18:37
 Operator : JC/MD
 Sample : N5558-06 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20221109

Quant Time: Nov 11 00:38:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	183977	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78702	51.327	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.660%	
35) Dibromofluoromethane	5.385	113	61091	46.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.100%	
50) Toluene-d8	8.653	98	225831	49.489	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.980%	
62) 4-Bromofluorobenzene	11.079	95	87367	49.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.020%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	689	0.541	ug/l	98
4) Vinyl Chloride	1.374	62	14713	12.420	ug/l	95
20) Methylene Chloride	2.800	84	1033	0.808	ug/l #	78
21) trans-1,2-Dichloroethene	3.105	96	435	0.361	ug/l #	90
27) cis-1,2-Dichloroethene	4.489	96	112422	80.737	ug/l	90
44) Trichloroethene	7.129	130	17681	12.145	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032737.D
Acq On : 10 Nov 2022 18:37
Operator : JC/MD
Sample : N5558-06 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20221109

Quant Time: Nov 11 00:38:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

