

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032738.D
 Acq On : 10 Nov 2022 19:00
 Operator : JC/MD
 Sample : N5558-05 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 45BR-20221109

Quant Time: Nov 11 00:38:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75456	50.575	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.160%
35) Dibromofluoromethane	5.385	113	59115	46.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.653	98	219405	49.667	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	83165	48.195	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.380%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	844	0.732	ug/l #	73
20) Methylene Chloride	2.788	84	1289	1.037	ug/l #	92
27) cis-1,2-Dichloroethene	4.489	96	52200	38.528	ug/l	89
44) Trichloroethene	7.129	130	225071	159.702	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032738.D
Acq On : 10 Nov 2022 19:00
Operator : JC/MD
Sample : N5558-05 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
45BR-20221109

Quant Time: Nov 11 00:38:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

