

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032739.D
 Acq On : 10 Nov 2022 19:23
 Operator : JC/MD
 Sample : N5565-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 221109040-01-TRIP-BLANK

Quant Time: Nov 11 00:38:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	168325	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73327	51.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.520%
35) Dibromofluoromethane	5.391	113	55833	46.501	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.647	98	207223	49.634	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.260%
62) 4-Bromofluorobenzene	11.079	95	78336	48.033	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.060%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4937	9.496	ug/l	99
30) Chloroform	5.099	83	2016	0.897	ug/l	99
44) Trichloroethene	7.129	130	690	0.518	ug/l	88
47) Bromodichloromethane	7.830	83	433	0.245	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032739.D
Acq On : 10 Nov 2022 19:23
Operator : JC/MD
Sample : N5565-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
221109040-01-TRIP-BLANK

Quant Time: Nov 11 00:38:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

