

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032751.D
 Acq On : 11 Nov 2022 02:31
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Nov 11 05:34:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 11 05:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	12130	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	84756	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	10.055	117	87449	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	35262	30.133	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 100.433%	
60) 4-Bromofluorobenzene	11.079	95	47114	28.165	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 93.900%	
63) Toluene-d8	8.653	98	89485	30.321	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 101.067%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032751.D
Acq On : 11 Nov 2022 02:31
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Nov 11 05:34:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 11 05:24:06 2022
Response via : Initial Calibration

