

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
 Data File : VX032753.D
 Acq On : 11 Nov 2022 04:04
 Operator : JC/MD
 Sample : VX1110WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1110WBL02

Quant Time: Nov 11 05:34:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 11 05:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	11124	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	77275	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	82185	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	32345	30.140	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.467%
60) 4-Bromofluorobenzene	11.079	95	42982	27.341	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.133%
63) Toluene-d8	8.653	98	82704	29.819	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111022\
Data File : VX032753.D
Acq On : 11 Nov 2022 04:04
Operator : JC/MD
Sample : VX1110WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1110WBL02

Quant Time: Nov 11 05:34:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 11 05:24:06 2022
Response via : Initial Calibration

