

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038726.D
 Acq On : 10 Nov 2023 14:15
 Operator : JC/MD
 Sample : 05241-13DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-24-20231102DL

Quant Time: Nov 11 01:53:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	255099	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	424094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	445281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	256596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149659	47.123	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.240%
35) Dibromofluoromethane	5.385	113	140195	45.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.800%
50) Toluene-d8	8.647	98	507186	46.403	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	216434	51.073	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.140%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	15784	6.630	ug/l	# 76
20) Methylene Chloride	2.788	84	2230	0.827	ug/l	85
27) cis-1,2-Dichloroethene	4.489	96	2972	0.993	ug/l	83
44) Trichloroethene	7.129	130	128856	37.896	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038726.D
Acq On : 10 Nov 2023 14:15
Operator : JC/MD
Sample : 05241-13DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-24-20231102DL

Quant Time: Nov 11 01:53:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

