

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038727.D
 Acq On : 10 Nov 2023 14:38
 Operator : JC/MD
 Sample : 05223-23DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-FD-2-20231031DL

Quant Time: Nov 11 01:54:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262877	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	437713	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	448045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	253744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	153908	47.027	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.060%		
35) Dibromofluoromethane	5.385	113	144007	45.680	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.360%		
50) Toluene-d8	8.647	98	511569	45.348	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	90.700%#		
62) 4-Bromofluorobenzene	11.079	95	218768	50.018	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.040%		
Target Compounds						
32) 1,1,1-Trichloroethane	5.385	97	6727	1.487	ug/l	Qvalue 96
44) Trichloroethene	7.129	130	163386	46.556	ug/l	88
64) Tetrachloroethene	9.281	164	1525	0.455	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038727.D
Acq On : 10 Nov 2023 14:38
Operator : JC/MD
Sample : 05223-23DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-FD-2-20231031DL

Quant Time: Nov 11 01:54:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

