

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038737.D
 Acq On : 10 Nov 2023 18:29
 Operator : JC/MD
 Sample : 05358-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D7-20231108

Quant Time: Nov 11 01:58:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257495	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	433091	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	449841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	254508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149722	46.705	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.400%		
35) Dibromofluoromethane	5.385	113	143361	45.960	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.920%		
50) Toluene-d8	8.647	98	512617	45.926	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	91.860%#		
62) 4-Bromofluorobenzene	11.079	95	216747	50.085	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.160%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	833	0.312	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.331	101	68618	27.358	ug/l	89
12) 1,1-Dichloroethene	2.325	96	18149	7.552	ug/l	81
24) 1,1-Dichloroethane	3.617	63	6336	1.476	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	9586	3.173	ug/l	77
30) Chloroform	5.099	83	5924	1.176	ug/l	94
38) Carbon Tetrachloride	5.678	117	6147	1.419	ug/l #	81
44) Trichloroethene	7.129	130	4541536	1307.890	ug/l	90
55) 1,1,2-Trichloroethane	9.159	97	2649	0.854	ug/l	92
64) Tetrachloroethene	9.275	164	11129	3.305	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038737.D
Acq On : 10 Nov 2023 18:29
Operator : JC/MD
Sample : 05358-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D7-20231108

Quant Time: Nov 11 01:58:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

