

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038738.D
 Acq On : 10 Nov 2023 18:52
 Operator : JC/MD
 Sample : 05358-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20231108

Quant Time: Nov 11 01:58:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	259176	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	431939	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	443724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150194	46.548	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.100%
35) Dibromofluoromethane	5.385	113	141711	45.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.100%
50) Toluene-d8	8.647	98	512727	46.058	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.120%
62) 4-Bromofluorobenzene	11.079	95	215100	49.837	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.680%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	2850	1.059	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.331	101	31264	12.384	ug/l	90
12) 1,1-Dichloroethene	2.319	96	6597	2.727	ug/l #	76
20) Methylene Chloride	2.794	84	1718	0.627	ug/l #	73
24) 1,1-Dichloroethane	3.611	63	2971	0.688	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	6475	2.129	ug/l	74
30) Chloroform	5.099	83	5403	1.066	ug/l	97
38) Carbon Tetrachloride	5.678	117	3385	0.784	ug/l	95
44) Trichloroethene	7.123	130	699894	202.096	ug/l	92
55) 1,1,2-Trichloroethane	9.153	97	1426	0.461	ug/l #	91
64) Tetrachloroethene	9.275	164	2041	0.614	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038738.D
Acq On : 10 Nov 2023 18:52
Operator : JC/MD
Sample : 05358-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20231108

Quant Time: Nov 11 01:58:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

