

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038740.D
 Acq On : 10 Nov 2023 19:38
 Operator : JC/MD
 Sample : 05351-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-241

Quant Time: Nov 11 01:59:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	260763	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	436221	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	462575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	258914	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	153692	47.342	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.680%		
35) Dibromofluoromethane	5.391	113	143665	45.727	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.460%		
50) Toluene-d8	8.647	98	532311	47.348	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.700%		
62) 4-Bromofluorobenzene	11.079	95	223776	51.338	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.680%		
Target Compounds						
						Qvalue
16) Acetone	2.379	43	45661	33.397	ug/l	# 88
20) Methylene Chloride	2.788	84	7697	2.793	ug/l	# 82
37) Ethyl Acetate	4.739	43	4351	1.092	ug/l	100
43) Isopropyl Acetate	6.336	43	147401	24.062	ug/l	# 91
44) Trichloroethene	7.129	130	6481	1.853	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038740.D
Acq On : 10 Nov 2023 19:38
Operator : JC/MD
Sample : 05351-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-241

Quant Time: Nov 11 01:59:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

