

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038747.D
 Acq On : 11 Nov 2023 00:31
 Operator : JC/MD
 Sample : VX1110WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1110WBL02

Quant Time: Nov 11 02:02:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	266253	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	447036	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	461028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	262489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152603	46.037	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.080%
35) Dibromofluoromethane	5.385	113	143545	44.584	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.160%
50) Toluene-d8	8.647	98	527389	45.775	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.560%#
62) 4-Bromofluorobenzene	11.079	95	224149	50.179	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038747.D
Acq On : 11 Nov 2023 00:31
Operator : JC/MD
Sample : VX1110WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1110WBL02

Quant Time: Nov 11 02:02:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

