

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038752.D
 Acq On : 11 Nov 2023 02:28
 Operator : JC/MD
 Sample : PB157051TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB157051TB

Quant Time: Nov 11 05:07:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	273909	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	463084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	479230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158818	46.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.140%
35) Dibromofluoromethane	5.385	113	151482	45.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.840%
50) Toluene-d8	8.647	98	549639	46.053	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.100%
62) 4-Bromofluorobenzene	11.079	95	229683	49.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.280%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	58146	40.487	ug/l	93
20) Methylene Chloride	2.788	84	12636	4.365	ug/l #	79
37) Ethyl Acetate	4.733	43	4326	1.023	ug/l	95
43) Isopropyl Acetate	6.336	43	166328	25.577	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038752.D
Acq On : 11 Nov 2023 02:28
Operator : JC/MD
Sample : PB157051TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB157051TB

Quant Time: Nov 11 05:07:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

