

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038753.D
 Acq On : 11 Nov 2023 02:51
 Operator : JC/MD
 Sample : 05350-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D1-20231107

Quant Time: Nov 11 05:07:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270325	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	449088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	468428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156730	46.570	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.140%		
35) Dibromofluoromethane	5.391	113	145989	45.135	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.280%		
50) Toluene-d8	8.647	98	535234	46.244	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.480%		
62) 4-Bromofluorobenzene	11.079	95	228664	50.956	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.920%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.337	101	2659	1.010	ug/l	Qvalue 87
12) 1,1-Dichloroethene	2.325	96	3104	1.230	ug/l	# 69
27) cis-1,2-Dichloroethene	4.495	96	3271	1.031	ug/l	73
38) Carbon Tetrachloride	5.684	117	5825	1.297	ug/l	96
44) Trichloroethene	7.129	130	363080	100.837	ug/l	91
55) 1,1,2-Trichloroethane	9.159	97	1496	0.465	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038753.D
Acq On : 11 Nov 2023 02:51
Operator : JC/MD
Sample : 05350-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D1-20231107

Quant Time: Nov 11 05:07:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

