

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038756.D
 Acq On : 11 Nov 2023 04:00
 Operator : JC/MD
 Sample : 05350-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT158S1-20231107

Quant Time: Nov 11 05:08:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	258605	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	424360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	441220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	256392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149232	46.352	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.700%
35) Dibromofluoromethane	5.385	113	139578	45.668	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	502181	45.916	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.840%#
62) 4-Bromofluorobenzene	11.079	95	216243	50.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.000%
Target Compounds						
						Qvalue
24) 1,1-Dichloroethane	3.611	63	8830	2.049	ug/l	99
44) Trichloroethene	7.135	130	1027	0.302	ug/l	73
52) Toluene	8.720	92	1817	0.254	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038756.D
Acq On : 11 Nov 2023 04:00
Operator : JC/MD
Sample : 05350-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT158S1-20231107

Quant Time: Nov 11 05:08:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

