

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038757.D
 Acq On : 11 Nov 2023 04:23
 Operator : JC/MD
 Sample : 05350-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT14911-20231107

Quant Time: Nov 11 05:08:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	254123	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	423675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	447062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	259087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147487	46.618	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.240%
35) Dibromofluoromethane	5.391	113	139585	45.744	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.480%
50) Toluene-d8	8.647	98	504942	46.243	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.480%
62) 4-Bromofluorobenzene	11.079	95	217218	51.309	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.620%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2823	1.051	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038757.D
Acq On : 11 Nov 2023 04:23
Operator : JC/MD
Sample : 05350-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT149I1-20231107

Quant Time: Nov 11 05:08:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

