

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038758.D
 Acq On : 11 Nov 2023 04:46
 Operator : JC/MD
 Sample : 05350-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20231107

Quant Time: Nov 13 01:05:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	264850	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	443929	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	465062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	271740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154204	46.767	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	93.540%	
35) Dibromofluoromethane	5.385	113	144942	45.332	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	90.660%	
50) Toluene-d8	8.647	98	527185	46.078	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	92.160%	
62) 4-Bromofluorobenzene	11.079	95	227780	51.349	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	102.700%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	3906	1.420	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	61162	23.708	ug/l	89
12) 1,1-Dichloroethene	2.325	96	25696	10.396	ug/l	# 76
24) 1,1-Dichloroethane	3.617	63	5519	1.250	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	8384	2.698	ug/l	74
30) Chloroform	5.105	83	8951	1.728	ug/l	99
38) Carbon Tetrachloride	5.678	117	11070	2.493	ug/l	98
44) Trichloroethene	7.123	130	1283243	360.531	ug/l	92
55) 1,1,2-Trichloroethane	9.153	97	3080	0.969	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038758.D
Acq On : 11 Nov 2023 04:46
Operator : JC/MD
Sample : 05350-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20231107

Quant Time: Nov 13 01:05:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

