

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038761.D
 Acq On : 11 Nov 2023 05:55
 Operator : JC/MD
 Sample : 05357-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20231107

Quant Time: Nov 13 01:06:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	264118	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	437715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	461320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	264832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154804	47.079	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.160%
35) Dibromofluoromethane	5.385	113	144226	45.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.500%
50) Toluene-d8	8.647	98	523685	46.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.840%
62) 4-Bromofluorobenzene	11.079	95	224772	51.390	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.780%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12668	9.148	ug/l	# 89
20) Methylene Chloride	2.788	84	3220	1.153	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038761.D
Acq On : 11 Nov 2023 05:55
Operator : JC/MD
Sample : 05357-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20231107

Quant Time: Nov 13 01:06:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

