

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
 Data File : VX038764.D
 Acq On : 11 Nov 2023 07:04
 Operator : JC/MD
 Sample : 05350-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT149S1-20231107

Quant Time: Nov 13 01:07:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261382	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	430285	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	450592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151217	46.469	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.940%
35) Dibromofluoromethane	5.385	113	141793	45.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.500%
50) Toluene-d8	8.647	98	512689	46.232	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.460%
62) 4-Bromofluorobenzene	11.079	95	218135	50.734	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.460%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	3.111	73	84696	10.858	ug/l	95
42) 1,2-Dichloroethane	6.098	62	1932	0.483	ug/l	100
44) Trichloroethene	7.129	130	20018	5.802	ug/l	87
64) Tetrachloroethene	9.275	164	3499	1.037	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	2080	0.257	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111023\
Data File : VX038764.D
Acq On : 11 Nov 2023 07:04
Operator : JC/MD
Sample : 05350-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT149S1-20231107

Quant Time: Nov 13 01:07:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

