

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111025\
 Data File : VX048538.D
 Acq On : 10 Nov 2025 14:45
 Operator : JC/MD
 Sample : Q3584-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-2

Quant Time: Nov 11 00:47:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	164671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	253875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	270369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142954	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	131909	57.488	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.980%
35) Dibromofluoromethane	5.373	113	114044	59.352	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.700%
50) Toluene-d8	8.634	98	309532	51.651	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.300%
62) 4-Bromofluorobenzene	11.061	95	129291	57.892	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111025\
Data File : VX048538.D
Acq On : 10 Nov 2025 14:45
Operator : JC/MD
Sample : Q3584-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-2

Quant Time: Nov 11 00:47:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

