

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111025\
 Data File : VX048545.D
 Acq On : 10 Nov 2025 17:10
 Operator : JC/MD
 Sample : Q3584-07 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEEP-1

Quant Time: Nov 11 00:49:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	176852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	345569	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	366760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	194556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	121452	49.285	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.560%
35) Dibromofluoromethane	5.379	113	116434	44.517	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.040%
50) Toluene-d8	8.634	98	405839	49.752	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.500%
62) 4-Bromofluorobenzene	11.061	95	175384	57.693	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.380%
Target Compounds						
65) Chlorobenzene	10.061	112	6669	0.766	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111025\
Data File : VX048545.D
Acq On : 10 Nov 2025 17:10
Operator : JC/MD
Sample : Q3584-07 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEEP-1

Quant Time: Nov 11 00:49:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

