

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111121\
 Data File : VX025137.D
 Acq On : 11 Nov 2021 18:41
 Operator : JC/MD
 Sample : M4645-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CON02

Quant Time: Nov 12 05:08:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 12 05:08:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	195759	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	173522	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	81550	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	65697	49.676	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.360%
7) Chloroethane-d5	1.666	69	51855	68.797	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	137.600%#
11) 1,1-Dichloroethene-d2	2.306	63	81652	35.897	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.800%
21) 2-Butanone-d5	4.458	46	101591	101.651	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.650%
24) Chloroform-d	5.062	84	113506	48.787	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.580%
26) 1,2-Dichloroethane-d4	5.958	65	70114	49.718	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.440%
32) Benzene-d6	5.976	84	242313	51.165	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.320%
36) 1,2-Dichloropropane-d6	7.312	67	75561	52.252	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.500%
41) Toluene-d8	8.653	98	225436	49.846	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.700%
43) trans-1,3-Dichloroprop...	8.952	79	39002	49.684	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.360%
47) 2-Hexanone-d5	9.384	63	78978	101.707	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.710%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	103029	49.467	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.940%
66) 1,2-Dichlorobenzene-d4	12.323	152	83821	51.850	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111121\
Data File : VX025137.D
Acq On : 11 Nov 2021 18:41
Operator : JC/MD
Sample : M4645-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CON02

Quant Time: Nov 12 05:08:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 12 05:08:01 2021
Response via : Initial Calibration

