

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111121\
 Data File : VX025138.D
 Acq On : 11 Nov 2021 19:04
 Operator : JC/MD
 Sample : M4645-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CON00

Quant Time: Nov 12 05:09:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 12 05:08:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	192003	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	171372	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	80632	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	63797	49.183	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.360%
7) Chloroethane-d5	1.666	69	49207	66.561	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	133.120%#
11) 1,1-Dichloroethene-d2	2.306	63	79863	35.797	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.600%
21) 2-Butanone-d5	4.458	46	99940	101.955	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.960%
24) Chloroform-d	5.062	84	111875	49.026	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.060%
26) 1,2-Dichloroethane-d4	5.964	65	69783	50.451	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.900%
32) Benzene-d6	5.977	84	239732	51.255	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.500%
36) 1,2-Dichloropropane-d6	7.312	67	74314	52.035	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.060%
41) Toluene-d8	8.653	98	222696	49.857	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.720%
43) trans-1,3-Dichloroprop...	8.952	79	37378	48.213	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.420%
47) 2-Hexanone-d5	9.384	63	78332	102.140	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.140%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	101848	49.513	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.020%
66) 1,2-Dichlorobenzene-d4	12.323	152	81953	51.272	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111121\
Data File : VX025138.D
Acq On : 11 Nov 2021 19:04
Operator : JC/MD
Sample : M4645-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CON00

Quant Time: Nov 12 05:09:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 12 05:08:01 2021
Response via : Initial Calibration

