

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111122\
 Data File : VX032764.D
 Acq On : 11 Nov 2022 11:09
 Operator : JC/MD
 Sample : N5565-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 221109040-01-TRIP-BLANK

Quant Time: Nov 14 05:10:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 11 05:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	10891	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	78733	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	80513	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	33411	31.799	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.000%
60) 4-Bromofluorobenzene	11.079	95	42897	27.853	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.833%
63) Toluene-d8	8.653	98	82896	30.509	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.700%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	1323	9.043	ug/l #	54
25) Chloroform	5.105	83	1458	0.617	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111122\
Data File : VX032764.D
Acq On : 11 Nov 2022 11:09
Operator : JC/MD
Sample : N5565-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
221109040-01-TRIP-BLANK

Quant Time: Nov 14 05:10:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 11 05:24:06 2022
Response via : Initial Calibration

