

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111122\
 Data File : VX032767.D
 Acq On : 11 Nov 2022 12:19
 Operator : JC/MD
 Sample : N5562-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 14 05:11:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 11 05:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	10027	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	71572	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	76962	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	30396	31.422	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.733%
60) 4-Bromofluorobenzene	11.079	95	39625	26.916	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.733%
63) Toluene-d8	8.653	98	77039	29.661	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.867%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111122\
Data File : VX032767.D
Acq On : 11 Nov 2022 12:19
Operator : JC/MD
Sample : N5562-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 14 05:11:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 11 05:24:06 2022
Response via : Initial Calibration

