

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111124\
 Data File : VX043778.D
 Acq On : 11 Nov 2024 09:45
 Operator : JC/MD
 Sample : VX1111MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 12 00:51:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 12 00:51:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	340497	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	296520	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	136329	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	73408	40.469	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.940%
7) Chloroethane-d5	1.642	69	38097	48.860	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.720%
11) 1,1-Dichloroethene-d2	2.288	65	35806	42.759	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.520%
21) 2-Butanone-d5	4.459	46	107469	90.486	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.490%
24) Chloroform-d	5.044	84	160700	41.095	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.200%
26) 1,2-Dichloroethane-d4	5.946	65	104410	42.030	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.060%
32) Benzene-d6	5.964	84	340715	44.444	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.880%
36) 1,2-Dichloropropane-d6	7.300	67	102748	44.053	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.100%
41) Toluene-d8	8.641	98	307016	44.086	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.180%
43) trans-1,3-Dichloroprop...	8.946	79	45509	40.800	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.600%
47) 2-Hexanone-d5	9.385	63	89362	96.813	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.810%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	140982	47.111	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.220%
66) 1,2-Dichlorobenzene-d4	12.317	152	113575	48.083	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111124\
Data File : VX043778.D
Acq On : 11 Nov 2024 09:45
Operator : JC/MD
Sample : VX1111MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 12 00:51:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 00:51:02 2024
Response via : Initial Calibration

