

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111124\
 Data File : VX043779.D
 Acq On : 11 Nov 2024 10:08
 Operator : JC/MD
 Sample : VX1111WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 12 00:51:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 12 00:51:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	343326	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	297696	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	133244	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	86593	47.344	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.680%
7) Chloroethane-d5	1.642	69	46608	59.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	118.560%
11) 1,1-Dichloroethene-d2	2.294	65	39653	46.963	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.920%
21) 2-Butanone-d5	4.465	46	127183	106.202	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.200%
24) Chloroform-d	5.050	84	186828	47.383	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.760%
26) 1,2-Dichloroethane-d4	5.952	65	120217	47.994	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.980%
32) Benzene-d6	5.964	84	390376	50.720	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.440%
36) 1,2-Dichloropropane-d6	7.306	67	117587	50.216	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.440%
41) Toluene-d8	8.647	98	350620	50.148	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.300%
43) trans-1,3-Dichloroprop...	8.952	79	53631	47.892	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.780%
47) 2-Hexanone-d5	9.384	63	99403	107.266	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.270%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	167396	55.717	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.440%
66) 1,2-Dichlorobenzene-d4	12.317	152	129959	56.293	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111124\
Data File : VX043779.D
Acq On : 11 Nov 2024 10:08
Operator : JC/MD
Sample : VX1111WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 12 00:51:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 00:51:02 2024
Response via : Initial Calibration

