

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111125\
 Data File : VX048556.D
 Acq On : 11 Nov 2025 12:04
 Operator : JC/MD
 Sample : Q3584-07 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEEP-1

Quant Time: Nov 12 04:44:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	213710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	420027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	432332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	225901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	146403	49.163	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.320%
35) Dibromofluoromethane	5.373	113	138352	43.520	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.040%
50) Toluene-d8	8.635	98	495865	50.012	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.020%
62) 4-Bromofluorobenzene	11.061	95	207015	56.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.060%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	4576	8.110	ug/l #	91
16) Acetone	2.380	43	7007	4.971	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111125\
Data File : VX048556.D
Acq On : 11 Nov 2025 12:04
Operator : JC/MD
Sample : Q3584-07 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEEP-1

Quant Time: Nov 12 04:44:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

