

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111125\
 Data File : VX048558.D
 Acq On : 11 Nov 2025 12:45
 Operator : JC/MD
 Sample : Q3574-01 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 304641-01 Liquid

Quant Time: Nov 12 04:45:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	167862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	340201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	363054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	194663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	127438	54.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.960%
35) Dibromofluoromethane	5.379	113	115536	44.871	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.740%
50) Toluene-d8	8.635	98	413001	51.429	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.860%
62) 4-Bromofluorobenzene	11.061	95	176457	58.962	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.920%
Target Compounds						
						Qvalue
44) Trichloroethene	7.117	130	985	0.386	ug/l	98
84) 1,2,4-Trimethylbenzene	11.732	105	5565	0.479	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111125\
Data File : VX048558.D
Acq On : 11 Nov 2025 12:45
Operator : JC/MD
Sample : Q3574-01 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
304641-01 Liquid

Quant Time: Nov 12 04:45:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

