

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005902.D
 Acq On : 12 Nov 2018 16:01
 Operator : JC/MD
 Sample : J5912-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1TB01-181108

Quant Time: Nov 13 11:39:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	122578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	175504	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	157241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	77250	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	69693	49.46	ug/l	0.00
Spiked Amount			Recovery	=	98.92%	
35) Dibromofluoromethane	5.51	113	56039	46.95	ug/l	0.00
Spiked Amount			Recovery	=	93.90%	
50) Toluene-d8	8.71	98	206625	52.07	ug/l	0.00
Spiked Amount			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	11.13	95	71357	50.01	ug/l	0.00
Spiked Amount			Recovery	=	100.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005902.D
Acq On : 12 Nov 2018 16:01
Operator : JC/MD
Sample : J5912-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1TB01-181108

Quant Time: Nov 13 11:39:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

