

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005904.D
 Acq On : 12 Nov 2018 16:54
 Operator : JC/MD
 Sample : PB114704TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114704TB

Quant Time: Nov 13 05:58:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	116424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	170290	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	159321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	75641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	68721	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
35) Dibromofluoromethane	5.51	113	55647	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
50) Toluene-d8	8.71	98	213432	55.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.86%	
62) 4-Bromofluorobenzene	11.14	95	72523	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
Target Compounds						
16) Acetone	2.45	43	4835	6.709	ug/l	91
20) Methylene Chloride	2.86	84	4316	3.646	ug/l	94
43) Isopropyl Acetate	6.46	43	8644	2.763	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005904.D
Acq On : 12 Nov 2018 16:54
Operator : JC/MD
Sample : PB114704TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114704TB

Quant Time: Nov 13 05:58:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

