

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005906.D
 Acq On : 12 Nov 2018 17:47
 Operator : JC/MD
 Sample : J5899-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-0146

Quant Time: Nov 13 06:02:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	115723	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	165971	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	150470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	74135	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	68003	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	5.51	113	55497	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	8.71	98	199047	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
62) 4-Bromofluorobenzene	11.14	95	68278	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
Target Compounds						
16) Acetone	2.44	43	2419	3.377	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	1504	1.394	ug/l #	79
27) cis-1,2-Dichloroethene	4.60	96	18439	15.478	ug/l	84
65) Chlorobenzene	10.14	112	6762	2.243	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005906.D
Acq On : 12 Nov 2018 17:47
Operator : JC/MD
Sample : J5899-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-0146

Quant Time: Nov 13 06:02:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

